



Abbott



The Ins and Outs of Training, Competency and Educational Requirements

11 | October | 2018

Jane L. Smith MS MT(ASCP) SI, DLM
Technical Manager, Scientific Affairs
Rapid Diagnostics



Agenda

1 Training Requirements

2 Competency and Educational Requirements

3 Common Questions

4 Show and Tell



Training Requirements

Training is Different than Initial Competency

Competency is **NOT**
the same as
performance
evaluation, proficiency
testing or training

Who Approves Training?

Laboratory
Director/Technical
Consultant

- The vendor can help the technical consultant with the initial training, but the facility's technical consultant(s) would have to sign off on the training

CAP – How Long Do You need to Keep Training Records?

GEN.55450 Initial Training Phase II

```
graph TD; A[GEN.55450 Initial Training Phase II] --> B[There are records that all laboratory personnel have satisfactorily completed initial training on all instruments/methods applicable to their designated job.]; B --> C[NOTE: The records must cover all testing performed by each individual. Training records must be maintained for a minimum of two years (five years for transfusion medicine). After the initial two year (or five-year) period, records of successful ongoing competency assessment may be used to demonstrate compliance with this requirement.]; C --> D[Retraining must occur when problems are identified with personnel performance.];
```

There are records that all laboratory personnel have satisfactorily completed initial training on all instruments/methods applicable to their designated job.

NOTE: The records must cover all testing performed by each individual. Training records must be maintained for a minimum of two years (five years for transfusion medicine). After the initial two year (or five-year) period, records of successful ongoing competency assessment may be used to demonstrate compliance with this requirement.

Retraining must occur when problems are identified with personnel performance.

CAP POC.09500 PPT Training

There are records demonstrating that all providers have satisfactorily completed initial training on the performance of the specific tests performed. NOTE: Medical staff credentialing is not an acceptable record of training.

CAP, TJC and COLA Specimen Collection Training Pre-Analytical

There are records that all personnel collecting patient specimens have been trained in collection techniques and in the proper selection and use of equipment/supplies and are knowledgeable about the contents of the specimen collection procedures.

- NOTE: This applies to all personnel who work under a single CLIA license.

All types of specimen collection techniques (e.g. phlebotomy, capillary, arterial, in-dwelling line, phlebotomy during intravenous infusion), as well as non-blood specimens, must be included in the training in accord with the individuals' duties.

Specimen collection for TJC is done initially(training), and then assessed and documented every 2 years. HR.01.06.01 Assessing phlebotomy staff competency

COLA includes an initial training, 6 month competency first year, and every year.




Competency and Educational Requirements

Why are Competency Requirements Confusing?

CLIA
regulations
for
competency
assessment
have not
changed


- Vague language
- Misinterpretation
- Various related requirements are interspersed throughout the CLIA regulations
- Requirements are not the same amongst the different inspecting groups

Who needs a Competency Assessment?



Which staff members should be assessed by whom?

The Laboratory Director is not required to have competency assessment – BUT is responsible for all CLIA defined responsibilities. The qualifications of individuals assessing competency of waived testing personnel shall be determined by the laboratory director.



*In addition to the six required assessments for testing performed, must also assess for competency **based on their federal regulatory responsibilities**

Example Competency Assessment for TC, TS, and GS

Assures that
performance
specifications are
established or verified
for necessary tests

Example Competency Assessment for TC, TS, and GS

Enrollment in an approved HHS approved proficiency testing program for each test requiring proficiency testing (PT)?

How well does the laboratory perform PT?

Review of PT results

Example Competency Assessment for TC, TS, and GS

Ensure that a Quality Control (QC) program is in effect and is adequate for the laboratory's testing .

Example Competency Assessment for TC, TS, and GS

Resolves technical
problems and
insures remedial
actions are taken

Example Competency Assessment for TC, TS, and GS

Ensures patient test results are not reported until all corrective actions have been taken and the test system is functioning properly


Example Competency Assessment for TC, TS, and GS

Identifies training needs
and assures that each
individual performing tests
receives regular in-service
training and education
appropriate for the tests
they are perform

Example Competency Assessment for TC, TS, and GS

Evaluates the competency of the testing personnel and assure that all staff members maintain their competency to perform tests accurately, report results promptly, accurately and proficiently.

Who Can Perform the Assessments?



Peer Testing Personnel (TP) cannot be designated to perform competency assessment if they do not qualify as General Supervisor (GS), Technical Consultant (TC), Technical Supervisor (TS)

Who Can Perform Competency Assessments for High-Complexity Testing?

Section Director

Technical Supervisor

- Bachelor's degree and 4 years training or experience in high-complexity testing

General Supervisor
(Delegated in writing by Section Director)

- Associate degree and 2 years of high complexity testing training or expertise. For technologists who've been working in the lab longer, the regulation makes an exception for those previously qualified as a general supervisor under federal regulations on or before Feb. 28, 1992. Also, someone at least meeting the minimum qualifications for a general supervisor.

Who Can Perform Competency Assessments for Moderate-Complexity Testing?

Technical Consultant

- Bachelor's degree and 2 years of laboratory training or experience with non-waived testing. The experience should be in the designated specialty or subspecialty in which the testing takes place

Trained Nurses (Delegated by LD)

- Bachelor's degree and 2 years of training or experience with non-waived testing. In addition, the laboratory director must delegate this task to a nurse in writing beforehand

2 year-degree RN, anesthesia tech, respiratory tech etc. CANNOT assess moderate complexity testing, even if delegated by the Lab Director

Does CLIA Require Competency for Waived Testing?

NO COMPETENCY ASSESSMENT is required for personnel who only perform waived testing in a CLIA laboratory

Personnel involved in pre- or post-analytical processes are not required to undergo competency assessment

Waived testing personnel, non-testing pre/post analytic personnel & those not in regulatory positions aren't subject to competency assessment

CLIA Competency Assessment Policy


Annual CA is required for all technical, supervisory & testing personnel.

Current staff need CA before patient testing when new methods or instruments are added.

Various related requirements are interspersed throughout regulations.

New staff have CA semiannually.


Six elements are necessary for all who perform non-waived testing, for all tests performed.



CA must be documented.

Operator training prior to testing is critical & required.

Six Elements for Non-waived Testing



CAP Waived Testing Competency

It is not necessary to assess all 6 elements for each assessment event:
The **POC program may select which elements to assess.**
Selected elements of competency assessment include but are **not limited to the six elements** required for non-waived testing

A laboratory must evaluate and document the competency of all testing personnel for each test system

Any personnel whose work is part of the testing process (includes pre-analytical) CAP Qprobe QP174

Identify which test systems each person uses

Question for CAP email accred@cap.org

Hi Jane,

The inspectors instructions are correct, the competency assessment must be performed at each location, even if everything is identical.

Some helpful ideas might be using the same form for the assessment and indicating on the form the site. Also keeping the records electronically might help with the organization and allow easy access or a more proficient check to see if any elements or sites have not been assessed. One facility had testers listed along with the sites they performed testing at, and even included hyperlinks to the documentation at each location.

I hope this helps!

Sincerely,

Jean Hood

Team Lead Inspection Services

Laboratory Accreditation Program

----- Original Message -----

Hello CAP Technical Services,

Can you advise on this question? I would like to inquire more specifically about competency assessment for EPOC operators who work at multiple CLIA locations. We have a sizeable group of EPOC operators who work at the main hospital and satellite NICUs at three other CLIA locations run by us, but located within other regional hospitals. An inspector indicated that operators would have to complete and document annual competency assessments at all 4 locations each year, even though the tests are performed and overseen identically at all 4 locations under one shared set of written procedures. With the specifics of our situation in mind, have you had any experience with a customer faced with this challenge? Because it would be logistically challenging to pull off assessments of each individual at 4 locations, I am wondering if you have heard how of any strategies that any other labs have used to meet this requirement.

Best regards,

Jane

CAP Provider Performed Testing

Competency Assessment Required:


- PPT is performed under the same CLIA number as the laboratory, and
- The laboratory director is responsible for competency assessment of the physicians and midlevel practitioners.

Competency Assessment NOT Required


- PPT is performed (waived testing only) under the same CLIA number as the laboratory, and
- The institutional medical staff has established the competency of physicians and mid level practitioners through the credentialing process.

TJC states if PPT does not involve an instrument, waived PPT may use medical staff credentialing for training and competency. WT.03.01.01 EP6

TJC Waived Testing Competency



COLA Waived Competency Assessment





Common Questions

CAP Q&A

Do I have to observe EVERY user?

- If they are performing Non-waived testing, Yes.
- If they are performing Waived testing, observation is optional.

Do I have to do competency assessment on non-laboratory employees?

- If they are performing testing, they need to have a competency assessment.

If I assess the employee at the end of training, is this considered the first competency for a user?

- You cannot use training as the your initial competency assessment. Competency is a separate process, and should occur approximately 4-6 months after user begins testing.

COLA Q&A

Who should perform the competency assessment on the Technical Consultant?

- The Lab Director, but this can also be done for example, by another Technical Consultant within the same group.

Who evaluates competency of LD or Clinical Consultant?

- Competency assessment is not required for the Lab Director. The Lab Director responsibilities will be evaluated in detail at the time of survey.
- If the Clinical Consultant and the Lab Director are the same person, competency assessment is not required. If they are two different people, then competency assessment is required for the Clinical Consultant.
- This should be done by the Lab Director, and is simply a review to determine if the CLIA responsibilities of the position are being met.

Is a competency assessment needed for specimen collection and processing personnel?

- YES

COLA Q&A

Can the Lab Manager monitor test performance by personnel, if the Lab Manager's competency assessment is performed by the lab director?

- “Lab Manager” is not a CLIA defined position. The Lab Manager must have minimum qualifications of a TC, TS or GS.

Should the lab director sign all competency evaluations?

- No, not necessarily. The TC, TS, or GS can sign the competency evaluations. This should be defined in the competency assessment procedures.

COLA Q&A

If the TS/TC/GS also performs testing, who can/should perform the direct observation of patient testing and instrument maintenance at a small high complexity facility? The lab director does not perform any testing functions.

- In a high complexity lab, both a TS and a GS are required positions. If these are different people, then the answer is easy; they can directly observe each other.
- If the TS and GS are the same person, and the Lab Director does not perform testing you could have other competent high complexity testing personnel do the direct observation component of the competency assessment.
- The final competency evaluation should be reviewed by the Lab Director in this case.

COLA Q&A

What do you do in a physician's office setting where the testing personnel is only one person who also serves as the general supervisor, and the lab director/tech supervisor is offsite and doesn't actually do any of the testing?

- The TP/GS could plan to do a self-evaluation in the presence of the LD/TS – and gather and review the documentation together.

COLA Q&A

If the physician is doing PPM can PT serve as his competency?

- If this physician is the Lab Director, then this is acceptable. If the physician is not the Lab Director, but rather is just one of the testing personnel for PPM, then the PT can be part of the competency assessment – but does not alone satisfy the requirement for competency assessment.

If lab director is also the TC in a small lab and performs some testing, who evaluates the Director?

- It is not required for the Lab Director to undergo competency assessment for the positions that he/she holds, including testing personnel.

COLA Q&A

As a working Lab Manager (MLT) do I qualify as the TC?

- “Lab Manager” is not a CLIA defined position. But the requirements to be a Technical Consultant, which is a required position for a moderate complexity lab, is a minimum of a bachelor’s degree in a chemical, biological, physical, or medical technology plus a minimum of two years training or experience in non-waived testing. An MLT can qualify, but must have the bachelor’s degree.

COLA Q&A

Can an MLT manager evaluate MT testing personnel if the Director has delegated it?

- CLIA says that the person who evaluates and determines competency must have qualifications of a Technical Consultant in a moderate complexity lab – or the qualification of a Technical Supervisor or General Supervisor in a high complexity lab.
- If an MLT meets these, there would be no problem with that person evaluating competency on others. The MLT would have to meet the qualifications. The Technical Consultant and Technical Supervisor qualifications include a minimum of a bachelor's degree, as far as education.

COLA Q&A

If the Laboratory Director is to document the competency assessment for the General Supervisor, what suggestions do you have to validate and document this?

- The Lab Director in this case would need to take some responsibility for being familiar with the test performance, documentation, problem solving, and reporting. They could read through the procedures together and make sure that all steps in the testing process are done as prescribed in the procedure. This should be documented as any other competency.

Can I Combine Competency Assessments on Large Platforms or Test Systems?

Acceptable if no unique aspects, problems or procedures in the tests

Tests with unique aspects, problems or procedures within the same testing platform must be assessed separately to ensure that staff maintain their competency to report test results promptly, accurately and proficiently

AACC POC List Serve

Date: Sun, 29 Jan 2017 15:37:36 +0000
From: "Mann, Peggy" <pmann@UTMB.EDU>
To: "aacc-poct-div@aacclists.org" <aacc-poct-div@aacclists.org>
Cc: "Michelle.Reeves@uhealth.org" <Michelle.Reeves@uhealth.org>
Subject: Re: [Aacc-poct-div] provider performed microscopy
Message-ID: <824574EF1D992F42940CC2C14216D6C9CB0244FE@GRMBX4.utmb.edu>
Content-Type: text/plain; charset="us-ascii"

Hi Michelle,

We by NO means 'have PPMP under 'control'. I threw caution to the wind, however, and tried to respond to your questions:

<Can I get a brief overview of everyone's structure for this?>

Our non-hospital clinics have CLIA PPMPs held by an MD within the department.

The POC lab has set up the competency assessment via an online 'test' with images, case studies. Departments using those competency assessment onlines are given input to the images used, case studies created. The onlines are done through the health system's compliance online testing system which sort of makes the 'who took it' assessment 'trackable' but completing the circle of having all the elements met and knowing who ALL are using the microscope/documenting in EMR is a gap (worse yet, who is failing to document) and is not where it needs to be for 100 compliance. The observation element has to be performed/conducted by faculty within the department holding the CLIA PPMP (eg Derm, Fam Med, Pedi).

<We are suddenly getting HUGE push back from our providers about having to do training, direct observations, comp assessments, etc. Has anyone else encountered this?>

Yes although for us it's been more constant over last 15 years. Never 'suddenly' a problem.

<How did you overcome that?>

I don't feel we in POCT have overcome resistance. I don't feel the health system (particularly the medical staff upper echelon) has been supportive enough, has not built in as yet the accountability required on the side of the departments holding the CLIAs.

Good Luck in your endeavor and be sure to post if you 'solve this' within your organization, Peggy

Show and Tell



References

CAP Checklists 8/17/2016
TJC Standards 2017
COLA Criteria 2016

QUESTIONS?

Photos are for illustrative purposes only. Any person depicted in photos is a model. SAHIGHR-0078 vA



Abbott